



All India Institute of Medical Sciences, Bhubaneswar
At - Sijua (Patrapada), Post - Dumuduma, Bhubaneswar (Odisha) - 751 019

File No- STORE-112/8/2024-STPUR SEC

Dated: 11-06-2025

Corrigendum(e-Tender)-I

e-Tender Ref. No.: STORE-112/8/2024-STPUR SEC/03

Dated:21-04-2025

Name of the Tender : Supply and Installation & Commissioning of Surgical Drill at AIIMS, Bhubaneswar.

SL No	Existing Specification	To be read as		
1	4. MICRO SAWS: a) Sagittal Saw-Qty 01 Should run at CPM of 25000 cpm	4. MICRO SAWS: a) Sagittal Saw-Qty 01 Should run at CPM of 20000 cpm or More		
2	5. UNIVERSAL DRIVER-QTY 01 Should run at RPM of 1500.	5. UNIVERSAL DRIVER-QTY 01 Should run at RPM of 1200 or more		
3	6. CONNECTING CORD-Qty 03 Length of cord should be up to 12 ft.	6. Integrated Micro Motor with Cord, Qty-03 Nos Length of cord should be up to 9ft or more.		
4	4. Rasp Cable length minimum of 12 ft and autoclavable. A minimum of three cables and a two-foot switch should be supplied.	4. Rasp Integrated Micromotor with Cable length minimum of 9ft and autoclavable. A minimum of three Micromotor with Cord and a two-foot switch should be supplied.		
5	Closing date & Time of Submission	06-062025 @5:00PM	Closing date & Time of Submission	25-06-2025@5:00PM
6	Date & time of online Technical Bid opening.	11-06-2025 @3:00PM	Date & time of online Technical Bid opening.	30-06-2025 @3:00PM

Note-All other terms and conditions of mentioned in the Tender remain Unchanged.

Sr. Procurement-cum Store Officer (I/c)
AIIMS, Bhubaneswar