



All India Institute of Medical Sciences, Bhubaneswar
At - Sijua (Patrapada), Post - Dumuduma, Bhubaneswar (Odisha) - 751 019

File No- STORE-22/20/2025-STPUR SEC

Dated: 10-07-2025

Corrigendum(e-Tender)-I

e-Tender Ref. No.: STORE-22/20/2025-STPUR SEC/19

Dated:02-06-2025

Name of the e-Tender: Supply, Installation, Commissioning and Maintaining 02 Units of Fully Automatic Biochemistry Analyzers on Reagent Rental Basis for Dept. of Biochemistry at AIIMS, Bhubaneswar for a period of Five years from Letter of Award.


SL No	SI No as per mentioned in Tender Document	Existing Specification	To be read as
1	Technical Bid Point 1(C)	1) System: The analyzer should have provision of integration with similar analyzers in the same laboratory.	1) System: The analyzer may have provision of integration with similar analyzers in the same laboratory.
2	Technical Bid Point 6(C)	Reagent: On board reagent stability of at-least 15 days or higher	Reagent: On board reagent stability of at-least 7 days or higher.
3	Technical Bid Point 10	Sample Capacity: Minimum 80 or more samples loading capacity at single time with continuous random access loading facility, thereafter.	Sample Capacity: Minimum 60 or more samples loading capacity with continuous random access loading facility, thereafter.
4	Technical Bid Point 11	STAT facility: STAT prioritization for biochemistry and ISE should be available on the system without interrupting the routine run with at least 20 stat sample positions.	STAT facility: STAT prioritization for biochemistry and ISE should be available on the system without interrupting the routine run.
5	Technical Bid Point 13(C)	System should have automatic re- run, automatic reflex testing and have facility for continuous loading.	System should have automatic re-run, OR automatic reflex testing and have facility for continuous loading”.
6	Technical Bid Point 14(b)	The vendor has to supply FREE OF COST: Specialty Internal Quality Control material for parameters like ADA, HbA1c, hsCRP, CSF Protein, Urine Microalbumin (in adequate quantity for running two levels twice a day)	The vendor has to supply FREE OF COST: Specialty Internal Quality Control material for parameters like ADA, HbA1c, hsCRP, Ferritin , CSF Protein, Urine Microalbumin (in adequate quantity for running two levels twice a day)
7	Technical Bid Point 18	Information Tech: The vendor shall have to integrate the analyzer with the existing LIS (CDAC e-Sushrut G5) available in the hospital. The equipment should be provided with software which should have capacity of data management and auto validation features.	Information Tech: “The vendor shall integrate the analyzer with the existing LIS (CDAC e-Sushrut G5) available in the hospital. The equipment should be provided with software which should have capacity of data. management”.

8	Technical Bid Point 20	<p>The quoted model must be able to perform all the listed tests:</p> <p>The bidder must compulsorily quote prices for all parameters with reagents from a single manufacturer. Relaxation will be allowed for any three of the four listed parameters whose test load is less than 5000 Tests/year (Adenosine deaminase, CSF Protein, Creatine Kinase, Gamma Glutamyl Transpeptidase) The bidder has to quote the price of reagents from third party for these three parameters, which will be supplied to be performed as open channel.</p>	<p>The quoted model must be able to perform all the listed tests:</p> <p>The bidder must compulsorily quote prices for all parameters with reagents from a single manufacturer. Relaxation will be allowed for any three of the four listed parameters whose test load is less than 5000 Tests/year (Adenosine deaminase, CSF Protein, Creatine Kinase, Gamma Glutamyl Transpeptidase) The bidder has to quote the price of reagents from third party for these three parameters, which will be supplied to be performed as open channel.</p> <p>The bidder is permitted to quote third party reagent kits for the parameters of Ferritin and hsCRP, ONLY when such reagents are not available from the OEM.</p> <p>A formal declaration on OEM letterhead has to be enclosed.</p>
9	Point 2.7.6	Conditional warranty like mishandling, manufacturing defects etc. is not acceptable.	Conditional warranty like mishandling of the analyzers during transport , manufacturing defects etc. is not acceptable.

Schedule:

SNo.	Existing		To be Read as	
01	Last date and time of submission of e-tender	Dt. 11-07-2025, at 17:00 Hrs	Last date and time of submission of e-tender	Dt. 24-07-2025, at 17:00 Hrs
02	Date & time of opening of e-Tender	Dt. 16-07-2025, at 15:00 Hrs	Date & time of opening of e-Tender	Dt. 28-07-2025, at 15:00 Hrs

Note: All other Terms & Conditions and Specification mentioned in the Tender remain unchanged.


Sr. Procurement-cum Store Officer(I/c)
AIIMS, Bhubaneswar