



1. Bid Details

Bid Number: **GEM/2026/B/7512180**

Bid Status: Active	Bid Start Date / Time: 15-05-2026 15:00:52	Consignees / Reporting Officer / Delivery Location(S)
Quantity: 1	Bid End Date / Time: 21-07-2026 11:00:00	EMD: Required Track EMD
Bid Validity (From End Date): 180 (Days)	Bid Opening Date / Time: 21-07-2026 11:30:00	Average Turn Over of Last 3 Years: 7 Lakh (s) Experience with Gov. Required: Year (s)
Competent Authority Document: View		

Buyer Details

Name: Himanshu Nirmal Chandu	Ministry: Ministry Of Health And Family Welfare	Organisation: All India Institute Of Medical Sciences (Aiims)
Address: Himanshu Nirmal Chandu,Stokee_himanshu@Aiimsbhubaneswar.Edu.In,AIIMS BHUBANESWAR, AT-SIJUA, PATRAPADA, POST-DUMUDUMA, BHUBANESWAR,KHURDA,ODISHA,751019,India	Department: Department Of Health And Family Welfare	Office: Bhubaneswar

Corrigendum Details

Modified On: 2026-07-10 12:18:28	Bid extended to 2026-07-21 11:00:00
	Bid Opening Date : 2026-07-21 11:30:00
Modified On: 2026-06-08 15:31:01	Bid extended to 2026-06-15 15:00:00 (Auto Extension)
	Bid Opening Date : 2026-06-15 15:30:00 (Auto Extension)

[Cancel Bid](#) [Extend Bid](#) [Edit Terms](#) [Edit Pre Bid](#)



Bid/RA

Note:

1. Response to all representation is mandatory.
2. Response to all Representations should be published at least 72 hours before bid end date and time. Buyer will not be allowed to open bid without responding to all representation.
3. Buyer can create responses and publish the same only after closure of representation window.
4. If you wish to modify the existing response or add new content, update the text and click Save again to apply the changes.

Bid Number: GEM/2026/B/7512180

Published

Bid/RA Section	Seller Query/Representation	Buyer Response	Submitted On
Product/Service specification	You Asked: X-Ray Generator: X-Ray Screen size: 7" or more. Our offer: Kindly amend it as X-Ray Screen size: 3.5" or more for healthy and wider participation.	As this is a ICMR sponsored project purchase, the specification fixed for the equipment can't be changed	16-05-2026 05:50 PM
	You Asked: X-Ray Tube: Anode Heat capacity : 8000j Our offer:- Kindly amend it as Anode Heat capacity : 6000j for healthy and wider participation.	As this is a ICMR sponsored project purchase, the specification fixed for the equipment can't be changed	16-05-2026 05:50 PM
	You Asked: Quality Standards & Certification: Both detector and X-Ray machine should be ICMR certified. Our offer:- Kindly delete this point for healthy and wider participation.	As this is a ICMR sponsored project purchase, the specification fixed for the equipment can't be changed	16-05-2026 05:50 PM

Close



Bid/RA

- Note:**
- 1. Response to all representation is mandatory.
 - 2. Response to all Representations should be published at least 72 hours before bid end date and time. Buyer will not be allowed to open bid without responding to all representation.
 - 3. Buyer can create responses and publish the same only after closure of representation window.
 - 4. If you wish to modify the existing response or add new content, update the text and click Save again to apply the changes.

Bid Number: GEM/2026/B/7512180

Pending

Bid/RA Section	Seller Query/Representation	Buyer Response	Submitted On
Product/Service specification	You Asked: X-Ray Generator: X-Ray Screen size: 7" or more. Our offer: Kindly amend it as X-Ray Screen size: 3.5" or more for healthy and wider participation.	As this is a ICMR sponsored project purchase, the specification fixed for the equipment can't be changed View	16-05-2026 05:50 PM
	You Asked: X-Ray Tube: Anode Heat capacity : 8000j Our offer:- Kindly amend it as Anode Heat capacity : 6000j for healthy and wider participation.	As this is a ICMR sponsored project purchase, the specification fixed for the equipment can't be changed View	16-05-2026 05:50 PM
	You Asked: Quality Standards & Certification: Both detector and X-Ray machine should be ICMR certified. Our offer:- Kindly delete this point for healthy and wider participation.	As this is a ICMR sponsored project purchase, the specification fixed for the equipment can't be changed View	16-05-2026 05:50 PM

Publish Representations

Save

Close