



**All India Institute of Medical Sciences, Bhubaneswar
At - Sijua (Patrapada), Post - Dumuduma, Bhubaneswar (Odisha) - 751 019**

Reference No : J-11021(001)/2013/S&P(7)

Dated : 23TH Dec 2013

ITEM : AUTOMATIC TISSUE PROCESSOR

The following 02 (Two firms) mentioned below are qualified on technical bid evaluation and requested to participate in the price bid opening at 3.30 PM on 26th Dec 2013 at Academic block , AIIMS Bhubaneswar.

1. **M/S NOBEL ENTERPRISES, BERHAMPUR, GANJAM, ODISHA**
2. **M/s SANTOSH AGENCIES., BERHAMPUR, GANJAM, ODISHA**

The following 01 (ONE) firms are disqualified in the technical bid for the reason mentioned below


Venders: Reasons for rejection

1. M/s RATTAN SALES CORPORATION ,NEW DELHI

The picture in the brochure does not show a carousal type construction in fact shows linear type. The picture in the brochure does not show the 12 station. Tissue basket are made up stainless steel. Delayed start function up to 90 Hours only. Option for changing the basket is not available in case of long time power failure.

Service support detail not available. Quoted model is not installation in eastern zone & no local service facility is available.

EC Declaration towards safety standards is not available, European EC/USFDA Approved certification not available.


Store & Purchase officer (I/C)

Copy to :

- | | | |
|-------------------------|---|--------------------------------------|
| 1. PS to Director | : | for kind information to the Director |
| 2. PS to DD (Admin). | : | for information of DD (Admin). |
| 3. PS to Fin. Advisor | : | for information of FA. |
| 4. HOD, Dept. Concerned | : | for information Please. |
| 5. Accounts Officer. | : | for information Please. |
| 6. Admn. Officer | : | for information Please. |
| 7. Dr.Manish Goyal | : | for hoisting in the website. |

Website: www.aiimsbhubaneswar.edu.in