



All India Institute of Medical Sciences, Bhubaneswar
At - Sijua (Patrapada), Post - Dumuduma, Bhubaneswar (Odisha) - 751 019

Reference No : J-11048(020)/2013/S&P

Dated : 17th Dec 2013

ITEM : LAPAROSCIOY SURGERY SET WITH HIGH DEFINITION CAMERA WITH ACCESORIES

The following 02 (Two firms) mentioned below are qualified on technical bid evaluation and requested to participate in the price bid opening at 3.00 PM on 20th Dec 2013 at Academic block , AIIMS Bhubaneswar.

1. M/S VISHAL SURGICALS EQUIPMENT CO., SECUNDERABAD
2. M/s HEALTHWARE PVT. LTD., HYDERBAD

The following 02 (Two) firms are disqualified in the technical bid for the reason mentioned below

Venders: Reasons for rejection

1. M/s SUSHRUTRA SURGICO MEDICAL PRODUCTS, BHUBANESWAR

All the instruments are not form the same manufacturer, the recording system is only software of another company, suction and irrigation unit is only irrigation one.

2. M/s JINDAL ENTERPRISES, KOLKATA

The vendor has quoted for the Richard Wolf model No.550910R2 for the CCU and camera but demonstration was given on model no.5525 Endocam logic HD Camera which is a higher version of the quoted model. The monitor mentioned 26 NDS Radiance monitor which is not from the same manufacturer. List of users has not been provided


Store & Purchase officer (I/C)

Copy to :

- | | | |
|-------------------------|---|--------------------------------------|
| 1. PS to Director | : | for kind information to the Director |
| 2. PS to DD (Admin). | : | for information of DD (Admin). |
| 3. PS to Fin. Advisor | : | for information of FA. |
| 4. HOD, Dept. Concerned | : | for information Please. |
| 5. Accounts Officer. | : | for information Please. |
| 6. Admn. Officer | : | for information Please. |
| 7. Dr.Manish Goyal | : | for hoisting in the website. |

Website: www.aiimsbhubaneswar.edu.in