



All India Institute of Medical Sciences, Bhubaneswar
At - Sijua (Patrapada), Post - Dumuduma, Bhubaneswar (Odisha) - 751 019

Reference No : J-11048(004)/2013/S&P

Dated : 5th Sep 2013

ITEM: ETO STERELIZER

The following 02 (Two firms) mentioned below are qualified on technical bid evaluation and requested to participate in the price bid opening at 4.30 PM on 6th Sep 2013 at Academic block, AIIMS Bhubaneswar.

1. M/s AGAMA SERVICES PVT. LTD – HYDERABAD-500 016
2. M/s 3M INDIA LIMITED - KOLKATA – 700 017

The following 03 (Three) firms are disqualified in the technical bid for the reason mentioned below

Venders: Reasons for rejection

1. M/s RELIANCE INSTRUMENTS CO.-CHENNAI
Indigenous cartridge, not EPA certified
2. M/s TAXUS MEDITECH
Indigenous cartridge, not EPA certified
3. M/s VISION CARE – BHUBANESWAR
No technical details / catalog provided. Self-declaration on EPA certification of cartridge not present.


Store & Purchase officer (I/C)

Copy to :

- | | | |
|---------------------------|---|--------------------------------------|
| 1. PS to Director | : | for kind information to the Director |
| 2. PS to DD (Admin). | : | for information of DD (Admin). |
| 3. PS to Fin. Advisor | : | for information of FA. |
| 4. Medical Superintendent | : | for information Please. |
| 5. Dr.Manish Goyal | : | for hoisting in the website. |