



All India Institute of Medical Sciences, Bhubaneswar
At - Sijua (Patrapada), Post - Dumuduma, Bhubaneswar (Odisha) - 751 019

Reference No : J-11048(010)/2013/S&P

Dated : 27th Aug 2013

ITEM: ELECTROCAUTERY

The following 3 (Three firms) mentioned below are qualified on technical bid evaluation and requested to participate in the price bid opening at 12.30 PM on 30th Aug 2013 at Academic block , AIIMS Bhubaneswar.

1. MEDILAB & CO. – KOLKATA-29
2. RABINDRA SURGICALS PVT. LTD. – BBSR- 20
3. VISHAL SURGICALS EQUIP.CO.- SECUNDERABAD-03

The following 04 (Four) firms are disqualified in the technical bid for the reason mentioned below

Venders: Reasons for rejection

1. TOSHBRO MEDICALS PVT. LTD. – KOLKATA – 700 017
No manufacturer and supplier ISO Certification
No-IEC-60601-1-2, No-ICE 60601-2-2
2. ALAN ELECTONIC SYSTEMS PVT. LTD. – MUMBAI- 42
List of users with performance certificate – Not Satisfactory
3. DIVLABS (INDIA) – NEW DELHI-77
After Sales service at Bhubaneswar not committed for specifically if service will be provided outside BBSR undertaking to bear the cost not provided.
4. MEDIPLUS – BHUBANESWAR
APC Compatibility not available
CMC Not Clear.


Store & Purchase officer

Copy to :

- | | | |
|---------------------------|---|--------------------------------------|
| 1. PS to Director | : | for kind information to the Director |
| 2. PS to DD (Admin). | : | for information of DD (Admin). |
| 3. PS to Fin. Advisor | : | for information of FA. |
| 4. Medical Superintendent | : | for information Please. |
| 5. Dr.Manish Goyal | : | for hosting in the website. |