



All India Institute of Medical Sciences, Bhubaneswar
At - Sijua (Patrapada), Post - Dumuduma, Bhubaneswar (Odisha) - 751 019

Reference No. J-11030(003)/2013/S&P

Dated : 20th Aug 2013

ITEM : ICU VENTILATORS

The following 5 (Five firms) mentioned below are qualified on technical bid evaluation and requested to participate in the price bid opening at 3.00 PM on 22nd Aug 2013 at Academic block , AIIMS Bhubaneswar.

1. AIR LIQUIDE MEDICAL SUG. PVT. LTD. CHENNAI – 600 097
2. WIPRO GE HEALTH CARE PVT. LTD – KOLKATA- 700 071
3. SCHILLER HEALTHCARE INDIA PVT. LTD. – KOLKATA – 700 064
4. COVIDIEN HELATH CARE PVT. LTD. – GURGAON – 122 002
5. MAQUET MEDICAL INDIA PVT. LT. – KOLKATA – 700 029

The following firms 03 (Three) are disqualified in the technical bid for the reason mentioned below

Venders: Reasons for rejection

1. DRAGER MEDICAL INDIA PVT. LTD. – KOLKATA – 700 107
Screen size, trend time and oxygen sensor are not matching the specification
2. PHILIPS ELECTONICS INDIA PVT. LTD. – KOLKATA – 700 020
Screen size and pressure flow loop are not matching the specification
3. MEDION HEALTHCARE PVT. LTD. – MUMBAI – 400 012
No mention against specifications no.20 and 23
Oxygen sensor not as asked for.


Store & Purchase officer

Copy to :

- | | | |
|---------------------------|---|--------------------------------------|
| 1. PS to Director | : | for kind information to the Director |
| 2. PS to DD (Admin). | : | for information of DD (Admin). |
| 3. PS to Fin. Advisor | : | for information of FA. |
| 4. Medical Superintendent | : | for information Please. |
| 5. Dr.Manish Goyal | : | for hosting in the website. |