



All India Institute of Medical Sciences, Bhubaneswar
At - Sijua (Patrapada), Post - Dumuduma, Bhubaneswar (Odisha) - 751 019

To be uploaded in Website: www.aiimsbhubaneswar.edu.in (Tender)

Reference No: J-11029(052)/2020-21/S&P

Dated: 08/08/2020

Sub: Procurement of "Cobe Spectra- LRS ELP Dual Needle Set" for the Department of Transfusion Medicine and Blood Bank at AIIMS, BBSR on Proprietary basis - Inviting Comments thereon.

The Department of Transfusion Medicine and Blood Bank of AIIMS, Bhubaneswar has requested for Procurement of Cobe Spectra- LRS ELP Dual Needle Set from **M/s Digital Specialty's, Bhubaneswar authorized dealers of M/s Terumo Penpol Pvt.Ltd,**

The Notice is being uploaded for general information of prospective manufacturer/Authorized Distributor/Dealers to submit their objection/proposal/comments, if any, on proprietorship of the item.

In case the product of any manufacturer/Authorized distributor/Dealer conforms to the enclosed specifications, they may submit their proposal for the supply of the same along with the brochures, point by point compliance of the enclosed specifications along with all documentary evidence. One quotation of the product may also be submitted.

The objections/comments/proposal should be sent in sealed cover or below e-mail id to the Office of Sr. Procurement Cum Stores Officer, AIIMS, Bhubaneswar (Odisha) – 751019 or through E-mail to spo@aiimsbhubaneswar.edu.in/aso.jkbiswal@aiimsbhubaneswar.edu.in/stokey-abnish@aiimsbhubaneswar.edu.in so as to reach on or before **date: 15-08-2020** failing which it will be presumed that no other firm is interested to offer comments/protest/object and case will be decided on its merits.

The ref. no. – J-11029(052)/2020-21/S&P dt: 08-08-2020, due on dated:15 -08-2020 should be super-scripted on sealed envelope.

Enclosure:

1. Proprietary Article Certificate from Indenting Officer of Department of Transfusion Medicine and Blood Bank, AIIMS, Bhubaneswar. **Annexure-I**
2. PAC of M/s. **TERUMO PENOL, KOLKATA. Annexure-II.**
3. E-mail Price Quotation of :Cobe Spectra- LRS ELP Dual Needle Set". **Annexure-III.**

Copy to:

1. Indenting Officer, Department of Transfusion Medicine and Blood Bank : for information please
2. Accounts Officer : for information please
3. IT Cell : for hoisting in the website.

Sr. Procurement-cum-Stores Officer
AIIMS, Bhubaneswar

AIIMS, BHUBANESWAR

Proprietary Article Certificate
Valid for the Current Financial Year

09

File No. and Date Reference :		
1	Description of article	Cobe Spectra LRS ELP Dual Needle set
2	Forecast of quantity /annual requirement	50 kit
3	Approximate estimated value for above quantity	4,79,167/- INR (approximate)
4	Maker's name and address	M/S Terumo Penpol Pvt. Ltd.
5	Name(s) of authorized dealers/ stockists	Digital Specialty's shop No.-3, Block- D/1, rajendra Vihar, Forest Park, Bhubaneswar-751009
6	I approve the above purchase on PAC basis and certify that : - Note- Tick to retain only one out of (b), C-1) or (c-2) whichever is applicable and cross out others. Please do confirm (a) by ticking It - without which PAC certificate will be invalid.	
6 (a)	This is the only firm who is manufacturing / stocking this item. AND	Yes, Attached
6 (b)	A Similar article is not manufactured / sold by any other firm, which could be used in lieu OR	Yes
6 (c-1)	No other make/brand will be suitable for following tangible reasons (like OEM/ Warranty, spares.) : OR	
6 (c)	No other make/brand will be suitable for following intangible reasons (if PAC was also given in the last procurement cycle, please also bring out efforts made since then to locate more sources) : OR	
7	Reference of concurrence of finance wing to the proposal :	
History of PAC Purchase of this item for past three years may be given below : NA		
Name of the Supplier		
Order/ Tender Reference & Date	Quantity Ordered	Basic Rate on Order (Rs.)
		Adverse Performance Reported if Any

Signature of Approving Authority.....

Date..... Designation of Officer.....

Signature of the Indenting Officer
Dr. Somnath Mukherjee
Associate Professor
AIIMS, Bhubaneswar-751019
Regd.No-57906(WB)

Saty - Prakash
20/12/2020

Dr. Satya Prakash
Assistant Professor
Dept. of Transfusion Medicine and Blood Bank
AIIMS, Bhubaneswar
MCI Regd. No.MP- 7638

Specification of Plateletpheresis Kit

- I. Kit should be compatible with Cobe spectra apheresis system.
- II. Kit must have option for concurrent multicomponent (Plasma and Platelet) collection through apheresis in separate bags.
- III. Kit must have sensor for access and return pressure detection.
- IV. Kit should have leukoreduction facility to reduce white blood cells content upto $<1 \times 10^6$.
- V. Kit should have pre-fitted 16G needle along with needle injury protector for apheresis procedure.
- VI. Kit should be sterilised prior with either Ethylene oxide or gamma irradiated.
- VII. Kit should be made up of biocompatible material.
- VIII. Platelet container of the kit should be made with suitable plasticizer to extend the shelf life of platelets bag for 5 days.
- IX. Plasma bag should be made up of material suitable for storage of plasma component in -80°C deep freezer for at least one year.
- X. Kit should not have any kink or obstruction in the tubing.
- XI. Kit should have chamber for collect concentration monitor and facility to detect any red cell contamination.
- XII. Shelf life of the kit should be at least one year at the time of delivery.
- XIII. Each Kit should be supplied along with 500ml ACD-A anti-coagulant.

Satya Prakash
14/07/2020

Dr. Satya Prakash
Assistant Professor
Department of Transfusion Medicine and Blood Bank
AIIMS, Bhubaneswar
MCI Regd. No. MP- 7638

Somnath Mukherjee
14/07/20

Dr. Somnath Mukherjee
Associate Professor
AIIMS, Bhubaneswar-751019
Regd. No. 57906(WB)

TERUMO PENPOL®

TERUMO PENPOL Private Limited
(Formerly Known as TERUMO PENPOL Limited)
East Zone Sales & Service Center :
Manasa Terrace
P-787, Lake Town, Block A, Kolkata-700 089
Phone : 033-40684092, E-mail : tpikolkata@terumopenpol.com

CIN No. U33112KL 1985PTC004531
website : www.terumopenpol.com
Corp. Mail ID : info@terumopenpol.com

PROPRIETARY CERTIFICATE

This is to certify that "consumables and disposables for the COBE Spectra" as listed below are being manufactured by TERUMO BCT which are specific to the equipment and proprietary in nature.

Product Code	Product's Description
70300	LRS ELP Dual Needle Set
70500	Therapeutic Plasma Exchange Set
70600	White Blood Cell Set
70700	Cobe Spectra RBCX Set

For TERUMO PENPOL Private Limited

SANJOY ROY CHOWDHURY
Regional Manager (East)
Mobile Number : 9830406866



Place : Kolkata

Date : October 9, 2018

Somnath Mukherjee
Dr. Somnath Mukherjee,
Associate Professor
AIIMS, Bhubaneswar-751017
Regd. No-57075/MCI

Satya Prakash
20/10/18
Dr. Satya Prakash
Assistant Professor
Dept. of Transfusion Medicine and Blood Bank
AIIMS, Bhubaneswar
MCI Regd. No. MP- 7638

Registered Office : P.B. No. 6105, I-2, Jawahar Nagar, Thiruvananthapuram - 695 003, Kerala, India
Ph No. : 0471-3015500/01, Fax : 0471-2721519
Central Marketing Office : No. 9, Padmanabha View, Gandhi Nagar, Second Main Road, Adyar, Chennai-600020
E-mail : tpimarketing@terumopenpol.com



05

Dr. Satya Prakash <tranmed_prakash@aiimsbhubaneswar.edu.in>

Quotation for Cobe Spectra LRS ELP Dual Needle Set

1 message

Chowdhury, Sanjoy <Sanjoy.Chowdhury@terumobct.com>

Tue, Jul 14, 2020 at 12:54 PM

To: Somnath Mukherjee <tranmed_somnath@aiimsbhubaneswar.edu.in>

Cc: "Dr. Satya Prakash" <tranmed_prakash@aiimsbhubaneswar.edu.in>, "MR. J. K. BISWAL"

<aso.jkbiswal@aiimsbhubaneswar.edu.in>, Finance Advisor <fa@aiimsbhubaneswar.edu.in>, Assistant Store Officer <aso@aiimsbhubaneswar.edu.in>, "Ghous, Gholam" <Gholam.Ghous@terumobct.com>, Deputy Director Administration <dda@aiimsbhubaneswar.edu.in>

Dear Sir,

Our present price for the subject kit=Rs.8,500/ + GST (12%). Our kits are locally available in Digital Specialty's, Shop No.3, Block-D/1, Rajendra Vihar, Forest Park, Bhubaneswar-751009, Telephone: 94370 74188, 70083-45890 and 0674-259-5821, E-Mail: digitalspecialtys@rediffmail.com

Since Cobe Spectra is now under sunseting phase, we do not have any current Purchase Order to submit herewith.

Thank you,

Yours faithfully,

Sanjoy Roy chowdhury

REGIONAL MANAGER, EAST

TERUMO PENPOL PRIVATE LIMITED

9830406866

sanjoy.chowdhury@terumobct.com

Sanjay Roy Chowdhury
Dr. Somnath Mukherjee
Associate Professor
AIIMS, Bhubaneswar-751019
Regd.No-57906(WA)

** Statement from Terumo BCT Inc. **

Confidential and/or privileged information may be contained in this message and any attachments. If you are not the intended recipient you should not copy or forward this message and should destroy it immediately. Any disclosure, copying, distribution and/or any other use of this communication by anyone other than the intended recipient is strictly prohibited and may be unlawful. The opinions, conclusions and other information in this message do not necessarily represent the views and/or opinions of the employer.

** End of statement **

Satya Prakash
Dr. Satya Prakash
Associate Professor
AIIMS, Bhubaneswar-751019
Regd.No-57906(WA)