




All India Institute of Medical Science, Bhubaneswar
At-Sijua (Patrapada), Post: Dumuduma, Bhubaneswar (Odisha)-751019



e- File No: STORE-54/45/2024-STPUR SEC

Date: 20-03-2025

Corrigendum-I

e-Tender Ref. No. : STORE-54/45/2024-STPUR SEC/49, Dated – 24-02-2025.

Name of the e-Tender : Supply, Installation & Commissioning of Inhaled Nitric Oxide Delivery and Monitoring System at AIIMS, Bhubaneswar.

SNo.	Existing in Technical Bid (Specification)	To be Read as	
01	3 (b) Capable of Automatic (preferably), Real-time, Semi- automatic and Manual NO delivery modes for use with adult, pediatric and Neonatal patients	Capable of automatic, real-time, semi-automatic, and manual NO delivery modes for use with adult, pediatric, and neonatal patients	
02	3 (h) Comprehensive alarms including that for cylinder pressure in both mbar and graphical.	Comprehensive alarms, including that for cylinder pressures in both mbar/mm of Hg and graphical.	
03	3 (m) Automatic Calibration & Remote Service support will be preferred	Automatic calibration & Remote service support will be required.	
04	Important: Point No- 04: Cylinders must be refillable in India with 800 PPM of NO and balanced N2 gas mixture. Cylinder refilling charges of the Nitric Oxide with pick-up to drop-back service at our Institute must be quoted clearly, which will remain in force for at least next 5 years	Cylinder must be refillable in India with 800 PPM of NO or more and a balanced N2 gas mixture. Cylinder refilling charges of the NO with a pick-up to drop-back service at our Institute must be quoted clearly in the financial Bid (Part-B), which will remain in force for at least the next 5 years.	
05	Important : Point No- 05: Bidders must quote the price for the INO system with consumables (NO delivery Kit compatible with SLE, Drager, Schiller ventilators), cost of cylinders (NO & Calibration Kit if any) and Warranty for 5 years, quote separately for CMC for another 5 years	The bidder has to supply one-time delivery of consumables (iNO delivery kit for neonates – 10 Nos. and calibration kit -1 Nos.) along with the Equipment and Sensor. The Equipment, sensor and calibration Kit should be covered under warranty for five years.	
06	Last date and time of submission of e-tender	Dt. 26-03-2025, at 05:00 PM	Last date and time of submission of e-tender Dt. 04-04-2025, at 05:00 PM
07	Date & time of opening of e-Tender	Dt. 28-03-2025, at 03:00 PM	Date & time of opening of e-Tender Dt. 07-04-2025, at 03:00 PM

Note: All other Terms & Conditions and Specification mentioned in the Tender remain unchanged.

Shubho
 20/3/25
 Sr. Procurement-cum Store Officer (I/C)
 AIIMS, Bhubaneswar