



All India Institute of Medical Sciences, Bhubaneswar
At - Sijua (Patrapada), Post - Dumuduma, Bhubaneswar (Odisha) - 751 019

To be uploaded in Website: www.aiimsbhubaneswar.edu.in (Tender) & CPPP

Reference No: J-11058(039)/2022-23/S&P

Dated: 26/09/2022

Subject: Procurement of 04 nos. Olympus Light-Guide Cable (accessories) and 01 no. Olympus Intra-Abdominal Insufflation Unit (UHI-4) for Laparoscopic Endoscopy Imaging System (Make: OLYMPUS) installed in the dept. of Urology at AIIMS, Bhubaneswar, on Proprietary basis - **Inviting Comments thereon.**

The Dept. of Urology at AIIMS, Bhubaneswar has requested for procurement of 04 nos. Olympus Light-Guide Cable (accessories) and 01 no. Olympus Intra-Abdominal Insufflation Unit (UHI-4) for Laparoscopic Endoscopy Imaging System (Make: OLYMPUS) on Proprietary basis from **M/s MediPlus, Kolkata (authorized dealer of M/s Olympus Medical Systems India Pvt. Ltd.)**

The Notice is being uploaded for general information of prospective manufacturer/ Authorized Distributor/ Dealers to submit their objection/ proposal/ comments, if any, on proprietorship of the item.

In case the product of any manufacturer/Authorized distributor/Dealer conforms to the enclosed specifications, they may submit their proposal for the supply of the same along with the brochures, point by point compliance of the enclosed specifications along with all documentary evidence. One quotation of the product may also be submitted.

The objections/comments/proposal should be sent in sealed cover to the **Office of Sr. Procurement Cum Stores Officer, AIIMS, Bhubaneswar (Odisha) – 751019** or through e-mail to spo@aiimsbhubaneswar.edu.in, aso.dkmaahana@aiimsbhubaneswar.edu.in and stokee_lovneet@aiimsbhubaneswar.edu.in, so as to reach on or before **date: 17-10-2022** failing which it will be presumed that no other firm is interested to offer comments/protest/object and case will be decided on its merits.

The ref. no. – **J-11058(039)/2022-23/S&P** dt: **26-09-2022**, due on dated: **17-10-2022** should be superscripted on sealed envelope.

Enclosure:

1. Proprietary Article Certificates from User Department at AIIMS, Bhubaneswar
2. Manufacturer's Proprietary Article Certificate-Cum-Compatibility Certificate dated 14-03-2022 & 15-03-2022
3. Manufacturer's Authorization letter dated 15-03-2022
4. Price Quotation Dated 02-05-2022

Copy to:

- | | | |
|-----------------------------------|---|------------------------------|
| 1. Addl. Professor & HOD, Urology | : | for kind information please |
| 2. Indenting Officer | : | for kind information please |
| 3. Accounts Officer | : | for kind information please |
| 4. IT Cell, AIIMS, Bhubaneswar | : | for hoisting in the website. |

Sr. Procurement-cum-Stores Officer

Sr. Procurement-cum-Stores Officer

AIIMS, BHUBANESWAR

Proprietary Article Certificate
Valid for the Current Financial Year

160

File No. and Date Reference :		
1	Description of article	Olympus Light Guide Cable.
2	Forecast of quantity /annual requirement	
3	Approximate estimated value for above quantity	Rs. 1,80,000/-
4	Maker's name and address	OLYMPUS
5	Name(s) of authorized dealers/ stockists	OLYMPUS
I approve the above purchase on PAC basis and certify that : - Note- Tick to retain only one out of (b), C-1) or (c-2) whichever is applicable and cross out others. Please do confirm (a) by ticking it - without which PAC certificate will be invalid.		
6 (a)	This is the only firm who is manufacturing / stocking this item. AND	
6 (b)	A Similar article is not manufactured / sold by any other firm, which could be used in lieu OR	
6 (c-1)	No other make/brand will be suitable for following tangible reasons (like OEM/ Warranty, spares.) : OR OEM - Compatibility with existing Olympus imaging System.	
6 (c)	No other make/brand will be suitable for following intangible reasons (if PAC was also given in the last procurement cycle, please also bring out efforts made since then to locate more sources) : OR	
7	Reference of concurrence of finance wing to the proposal :	Final available. Oce
History of PAC Purchase of this item for past three years may be given below :		
Name of the Supplier		
Order/ Tender Reference & Date	Quantity Ordered	Basic Rate on Order (Rs.)
HU/PCD/PMSS/AIIMS/28/18-19	10/5/2018	
	(02)	

Signature of Approving Authority.....

Date..... Designation of Officer.....

डा. आशुतोष बिस्वास
Dr. Ashutosh Biswas
कार्यकारी निर्देशक / Executive Director
एम्स, भुवनेश्वर
AIIMS, Bhubaneswar-751010

Signature of the Indenting Officer

Dr. Prasant Nayak
Additional Professor & Head
Department of Urology
AIIMS, Bhubaneswar-751010

AIIMS, BHUBANESWAR

Proprietary Article Certificate
Valid for the Current Financial Year

83

File No. and Date Reference :		
1	Description of article	Intox-abdominal Insufflation Unit.
2	Forecast of quantity /annual requirement	This will last for 5 years.
3	Approximate estimated value for above quantity	Rs. 5,00,000/-
4	Maker's name and address	OLYMPUS
5	Name(s) of authorized dealers/ stockists	OLYMPUS - MEDIPUS
I approve the above purchase on PAC basis and certify that : - Note- Tick to retain only one out of (b), C-1) or (c-2) whichever is applicable and cross out others. Please do confirm (a) by ticking it - without which PAC certificate will be invalid.		
6 (a)	This is the only firm who is manufacturing / stocking this item.	
6 (b)	A Similar article is not manufactured / sold by any other firm, which could be used in lieu OR	
6 (c-1)	No other make/brand will be suitable for following tangible reasons (like OEM/ Warranty, spares.) : OR com. compatibility with existing equipment.	
6 (c)	No other make/brand will be suitable for following intangible reasons (if PAC was also given in the last procurement cycle, please also bring out efforts made since then to locate more sources) : OR	
7	Reference of concurrence of finance wing to the proposal :	Fund available. <i>[Signature]</i>
History of PAC Purchase of this item for past three years may be given below :		
Name of the Supplier		
Order/ Tender Reference & Date	Quantity Ordered	Basic Rate on Order (Rs.)
ALL/PCD/PMSSV/AIIMS-11/28/18-19		
(10/5/2018)	01	

Signature of Approving Authority..... *[Signature]*

Date..... Designation of Officer.....

Signature of the Indenting Officer

[Signature]
11/2/2022
Dr. Prasant Nayak
Additional Professor & Head
Department of Urology
AIIMS, Bhubaneswar-75101



OLYMPUS MEDICAL SYSTEMS INDIA PRIVATE LIMITED

Ground Floor, Tower-C, SAS Tower, The Medicity Complex,
Sector – 38, Gurgaon 122001, Haryana, INDIA
Tel.:0124-4999191 Fax: 0124-4999190
Website: www.olympusmedical.co.in
CIN: U33110HR2009FTC039611

Date: 14th March 2022

Ref No: OMSI/21-22/3204

To,
The Director,
All India Institute of Medical Sciences,
Bhubaneswar, Sijua, Patrapada,
Orissa- 71019

Subject: - Proprietary Article Certificate

Dear Sir,

This is to certify & confirm that the below mentioned product are solely being manufactured by
M/s. Olympus Winter & IBE GMBH having their registered office at PO Box 701709, 22017
Hamburg Germany, Kuehnstrabe 61122045 Hamburg Germany.

Model No.	Model Description
WA03300A	WA03300A LIGHT-GUIDE CABLE, 2.8MM, 3M
WA03310A	WA03310A LIGHT-GUIDE CABLE, 4.25 MM, 3M

No other consumable from any other manufactures can be connected with existing Model- OTV-
S200/OTV-S300/OTV-S400/OTV-S190/OTV-SC2/ OTV-SC.

This model is Proprietary products of Olympus.

Yours faithfully,

For Olympus Medical Systems India Pvt. Ltd.,


Authorised Signatory



97

Date: 15th March 2022

Ref No: OMSI/21-22/3211

To,
The Director
All India Institute of Medical Sciences
Tatibandh, GE Road,

Raipur-492099 (CG)

Sub: Compatibility cum Proprietary Article Certificate for Olympus High Flow Insufflator (Model: UHI-4) for the department of Urology

Dear Sir,

We Olympus Medical Systems India Private Limited having our registered office at Ground Floor, Tower-C, SAS Tower, The Medicity Complex, Sector-38, Gurgaon - 122001, Haryana is subsidiary of M/s. Olympus Corporation of Asia Pacific Ltd of M/s Olympus corporation Japan.

We further state that M/s Olympus Corporation Japan having their office at 2951, Ishikawa- Cho, Hachioji-shi , Tokyo 192-8507, Japan is established, and reputable manufacturer of High Flow Insufflator (Model:UHI-4) and its accessories.

We hereby confirm that Olympus High Flow Insufflator (Model:UHI-4) is having 45L/Min flow rate and provides automatic smoke and mist evacuation in conjunction with existing and installed Olympus Thunderbeat System in department of Urology.

No other company is manufacturing the High Flow Insufflator (Model: UHI-4) which provides automatic smoke and mist evacuation with existing Installed Thunderbeat System.

Thanking you,

For Olympus Medical Systems India Pvt. Ltd.


Authorized Signatory

Rayal

OLYMPUS MEDICAL SYSTEMS INDIA PRIVATE LIMITED

Ground Floor, Tower-C, SAS Tower, The Medicity Complex, Sector - 38, Gurgaon 122001, Haryana, INDIA
Tel.: 0124-4999191, Fax: 0124-4999190 | Website : www.olympusmedical.co.in | CIN : U33110HR2009FTC039611

OLYMPUS

Date: 15th March 2022

Ref No: OMSI/21-22/3210

To,
The Director
All India Institute of Medical Sciences
Bhubaneswar (Odisha) - 751019

Sub: Procurement of "Olympus High Flow Insufflator (Model: UHI-4) for Department of Urology"

Dear Sir,

We Olympus Medical Systems India Private Limited having our registered office at Ground Floor, Tower-C, SAS Tower, The Medicity Complex, Sector-38, Gurgaon - 122001, Haryana is subsidiary of M/s. Olympus Corporation of Asia Pacific Ltd of M/s Olympus corporation Japan.

We further state that M/s Olympus Corporation Japan having their office at 2951, Ishikawa- Cho, Hachioji-shi, Tokyo 192-8507, Japan is established and reputable manufacturer of **High Flow Insufflator (Model: UHI-4) and its accessories.**

We do hereby authorize M/s. **MEDI PLUS** having their registered office at **574, Lewis Road, FL.No. B/108, 1st Floor, Madhab Nilaya Apartment, Bhubaneshwar (Odisha)-751014** to submit quote, related documents and subsequently negotiate and sign the contract with you against your requirement for the above goods manufactured by our principal company.

No company or firm or individual other than M/s. **MEDI PLUS** having their registered office at **574, Lewis Road, FL.No. B/108, 1st Floor, Madhab Nilaya Apartment, Bhubaneshwar (Odisha)-751014** are authorized to bid, negotiate, and conclude the contract in regard to this business against this specific subject matter.

The after Sales Services including Warranty, CMC would be handled directly Olympus Medical Systems India Pvt Ltd. / Authorized Service Partner. We also confirm that the price quoted by our distributor shall not exceed the price which we would have quoted directly.

This authorization is valid for above subject matter only.

Thanking you,

For Olympus Medical Systems India Pvt. Ltd.

Authorized Signatory



OLYMPUS MEDICAL SYSTEMS INDIA PRIVATE LIMITED

Ground Floor, Tower-C, SAS Tower, The Medicity Complex, Sector - 38, Gurgaon 122001, Haryana, INDIA
Tel.: 0124-4999191, Fax: 0124-4999190 | Website : www.olympusmedical.co.in | CIN : U33110HR2009FTC039611



Your Vision. Our Future

OLYMPUS MEDICAL SYSTEMS INDIA PVT LTD.
Megatherm Building, Unit B
7th Floor, Plot No. L1, Block GP, Sector V, KOLKATA 700091 India
Tel.: 033-66461400, Fax:..

STATE CODE : WB
GST NO : 19AABCO2131L1ZH

ACCESSORY / CONSUMABLE QUOTATION

Customer Code: 6005242 THE DIRECTOR ALL INDIA INSTITUTE OF MEDICAL, SCIENCES BHUBANESWAR SIJUA, DUMUDUMA, Khordha BHUBANESWAR 751019 India STATE CODE : OR GST No : 21AAAGA0127K1Z5	Quotation No. : 2077465 Date : 15.03.2022 Valid To : 25.03.2022
---	--

Kind Attn:-

S. No	Item Code	HSN/SAC Code	Item Desc	Qty	Unit price	Gross Amount	CGST		SGST/UTGST		IGST		CESS	Total Tax
							Rate	Amt	Rate	Amt	Rate	Amt		
1	WA03310A	85444999	WA03310A LIGHT-GUIDE CABLE, 4.25 MM, 3 M	2 PC	46,146.00	92,292.00		0.00		0.00	18.00	16,612.56	0.00	16,612.56
2	WA03300A	85444999	WA03300A LIGHT-GUIDE CABLE, 2.8MM, 3M, C	2 PC	29,912.00	59,824.00		0.00		0.00	18.00	10,768.32	0.00	10,768.32
Total					76,058.00	152,116.00	0.00		0.00		27,380.88		0.00	27,380.88
					Subtotal									152,116.00
					Total									179,496.88
					Round Off Amount									0.00
					Grand Total									179,496.88
INR : One Lakh Seventy Nine Thousand Four Hundred Ninety Six Eighty Eight Paise														
Terms & Conditions														

Corporate Office : Ground Floor, Tower-C, SAS Tower, The Medicity Complex, Sector- 38, GURGAON 122001, (Haryana), INDIA. CIN : U33110HR2009FTC039511

This is a system generated copy. Signature is not required.

Subject to GURGOAN jurisdiction
Website: www.olympusmedical.co.in

PAGE 1 of 2

OLYMPUS

Your Vision. Our Future

Payment Term: 100% advance along with P.O. All payments should be made by Demand Draft/Cheque in the name of "OLYMPUS MEDICAL SYSTEMS INDIA PRIVATE LIMITED ." payable at par at Delhi/Gurgaon(Haryana).

Delivery: Within 30 days from the date of receipt of amount PO and advance amount along with receipt of Entry Form from the Hospital.

Taxation Term: Taxes are applicable at the time of invoicing.

Transportation: Transportation and freight charge will be paid by the customer.

Cancellation: Once the order is placed on us, the buyer shall have no right in any circumstances to cancel the order without prior written consent of the company and subject to compliance by the buyer with such requirements, including requirement as to the payment of adequate compensation, as the company may impose as a condition of giving such consent.


Forms: The required forms according to the regulatory compliance of the state Government must be sent along with purchase order for commercial shipment of goods.

Warranty: Being a consumable item, no warranty will be given for this item.

Jurisdiction: Any dispute is subject to within the jurisdiction of Gurgaon Haryana state.

OLYMPUS MEDICAL SYSTEMS INDIA PVT LTD.

Customer's Acceptance


Name: Amar Dey
Title : Senior Engineer-Field Service
Date : 15.08.2022

Name :
Title :
Date :

A handwritten signature in blue ink.

**MEDi PLUS**

574, Lewis Road, FL.No.B/108, 1st Floor
Madhab Nilaya Apartment
Bhubaneswar-751014
Mob : 9437549485 / 9437036192
E-mail : mediplus.bbsr15@gmail.com

Ref: MP/DRAIIMS/BBSR/2021-22

Date: 17th March, 2022

To,
The Director
All India Institute of Medical Sciences,
Patrapada, Sijua,
Bhubaneswar.

Sub: Submission of quotation for High Flow Insufflators for the Department of Urology, AIIMS Bhubaneswar.

Respected Madam,

We are pleased to offer our lowest possible price for the above mention product. Hope the price will be accepted by your institution.

Offer

S.No	Model	Description	Qty	Tax	Prices as per qty	Prices with GST
1	UHI-4	UHI-4 HIGH FLOW INSUFFLATOR (220-240V)	1	12%	430,860.10	482,563.31
2	MAJ-1872	MAJ-1872 COMMUNICATION CABLE LONG	1	18%	76,988.00	90,845.84
3	MAJ-591	MAJ-591 SUCTION TUBE FOR UHI-2	2	18%	11,030.00	13,015.40
Total					518,878.10	586,424.55

OPTIONAL ITEMS					Prices as per qty	Prices with GST
1	MAJ-1939	MAJ-1939 SMOKE EVACUATION FOOT SWITCH	1	18%	59,425.00	70,121.50
2	WA58670A	WA58670A GAS PREHEATER GSH-2, 230 V, W/W	1	12%	127,505.40	142,806.05

Terms & Conditions :

1. Total Price are F.O.R Destination
2. Warranty- 5 years on Sr.no-1 (UHI-4) against any manufacturing defects
3. Warranty is not applicable on accessories (Sr. no 2 & 3)
4. CMC extra @ INR 8064+18% GST/year after 5 years warranty

Thanking You,

Yours Faithfully,
For Medi Plus,

Amit



Authorized Signatory

Deals in : Drugs, Laboratory Reagent, Surgical Goods,
Medical Equipments & Instrument, Hospital Furniture
& all Hospital Supply products.

GSTIN : 21ABKPM7659N1Z1

Drug Licence No. KH-17319-W, KH-17320-WC & KH-6000-WX

76

OLYMPUS

Your Vision, Our Future

45L HIGH FLOW INSUFFLATION UNIT

UHI-4

Advanced functionality for enhanced performance and reliability
in a demanding surgical environment



Days

UHI-4

Main Features

High Speed Insufflation

The UHI-4 offers an increased maximum flow rate of 45L per minute. In addition, a new display mode provides clear visualization of the pressure, flow rate and volume in real time for monitoring the status of the UHI-4.

Automatic Smoke Evacuation


An automatic smoke evacuation feature is enabled on the UHI-4 when it is coupled with a new or existing energy platform. This will help provide a clear and unobstructed view of the surgical field during laparoscopic procedures.

Adjustable Smoke Evacuation

In order to reduce the amount of CO₂ used during surgery, the UHI-4 allows for the smoke evacuation to be independently set on the front panel of the unit. The smoke evacuation can be toggled between a High and Low function.



New Cylinder Hose with Switch-over Valve




With the new CO₂ switch-over valve accessory, the surgeon can freely switch CO₂ tanks without interrupting long procedures, thus reducing the physical difficulty of exchanging tanks.

*To use this accessory, the new WM-NP2 mobile workstation, as well as the double cylinder holder (a WM-NP2 accessory) is required.


New Small Cavity Mode

The newly developed Small Cavity Mode allows for safer insufflation of cavities smaller than the abdomen.



Specifications

Dimensions (W x H x D)	370 x 120 x 300 mm
Weight	9.3 kg
Abdominal Pressure Control	3 to 25 mmHg
Flow Rate Settings	0.1 to 45 L/min
Cavity Mode	Normal/Small
Gas Supply From Wall Pipeline	Connectable
Alarm	Overpressure of abdominal cavity/Tube clogging/Insufficient supply of gas
Smoke Evacuation Function	Available (when connected to the below devices) • UES-40 Electrosurgical Unit • SonoSurg-G2 Sonosurg Generator
OR Integration	ENDOALPHA UCES-3 ver. 4
Standard Set	UHI-4 MAJ-590 Insufflation Tube MAJ-591 Suction Tube



Specifications, design and accessories are subject to change without any notice or obligation on the part of the manufacturer.

OLYMPUS

OLYMPUS MEDICAL SYSTEMS CORP.
Shinjuku Monolith, 2-3-1 Nishi-Shinjuku, Shinjuku-ku, Tokyo 163-0914, Japan

For a complete listing of sales and distribution locations visit:
www.olympus.com

Printed in Japan R0109E2-032012